

File no: E- 18867/M/561/2025-STR/9544

भारत सरकार
Government of India
स्वास्थ्य सेवा महानिदेशालय
Directorate General of Health Services
स्वास्थ्य एवं परिवार कल्याण मंत्रालय
Ministry of Health & Family Welfare
लेडी हार्डिंग मेडिकल कॉलेज एवं श्रीमती सुचेता कृपलानी अस्पताल
Lady Hardinge Medical College & Smt. Sucheta Kriplani Hospital
शहीद भगत सिंह मार्ग, नई दिल्ली- ११०००९
Shaheed Bhagat Singh Marg, New Delhi-110001
भंडार अनुभाग/Stores Section
Dated- 01.08.2025

PUBLIC NOTICE

The LHMC & Smt. S. K. Hospital intent to procure the following items/ stores under GFR-155 immediately;

Sr. No.	Name of Item	Description	Quantity Required	Warranty	Remarks
1	Cervix Thermoablator	Specifications attached	02 Nos.	Specifications attached	-

Quotation Format

Sr. No.	Name of item	Rate Offered (In Rs.)	Units	Packing Details	Quantity supplied to be	Amount Rupees In
	Offered Make/ Brand:					
	Offered model:					
Total						Rs.
GST Percentage						%
GST Amount						Rs.
Total Amount (included GST %)						Rs.
Rupees in Words: /- (Included GST)						
Note: 1. In case of discrepancy between rate offered (unit rate) and total amount written in figures, the rate offered (unit rate) shall prevail. 2. In case of discrepancy between total amount written in figures & words, the amount mentioned in the word shall prevail.						

Competent Suppliers/ Service Providers are requested to submit/ drop their quotations (as per the format) in the Tender Box kept in the Stores Section LHMC.

Apart from the above the quotations must bearing the following details as under:

Offered Warranty (if asked or applicable)

GST Number of the bidder:

Complete Mailing Address (with Pin Code):

Mobile Number:

Full Name of Owner (Please enclose the copy of PAN Card & Adhar Card- mandatorily submitted):

Name of Father/ Husband:

PAN No.:

Aadhar No.:

e-mail address:

Full Name of Signatory should be there in the quotation

Instructions-

1. The quotation should be addressed to the Chairperson Local Purchase Committee LHMC. The Subject of the quotation & Reference Serial Number of the Public Notice must be mentioned there on the Envelope.
2. The quotation & its envelope must bear Name, Contact Number and Address details of the bidder.
3. The demonstration of the quoted item must be provided if & when asked by the Institution, failing which the offer will be treated as rejected.
4. In order to safeguard the procurement, the Institute reserve the right to get a Performance Security of 10% or more through Bharat Kosh/ NRT Portal. Therefore, the bidders may be asked to submit the said Performance Security any time after submitting the quotations.
5. Authorization letter from the OEM & Warranty Undertaking etc against the said Public Notice are to be attached.
6. The supplier awarded with the Supply Order has to procure/ arrange the goods ONLY from the Authorized Distributor of the item & as a proof of the same the relevant document has to be submitted alongwith the supply of the goods failing which the payment will not be made.
7. The firms, have to submit the Drug License for alongwith the quotations wherever applicable.
8. The quotations, sent through email will not be entertained.
9. The quotations received after the prescribed dates will not be entertained in any condition.
10. The name & ID proof details (ADHAR Card Number, PAN Card Number) of the person coming to submit the quotations must be mentioned in the quotations.
11. In case of non-submission of any of the above details the quotation will be treated incomplete & rejected.
12. The quotation should be submitted between 10:30am to 12:30pm every working day. The last date of submission of quotation is 04/08/2025

Digitally signed by
AVANISH SINGH

(Avanish Singh) Date: 01-08-2025
Stores Officer 12:37:55
LHMC & Smt. S.K. Hospital, New Delhi

Specifications for Thermo-ablator

1. Portable Handheld thermal ablation device with rechargeable batteries.
2. Simple and easy to use, appropriate for all levels of care.
3. Probe tip temperatures controlled to reach 100 degrees Celsius
4. Probe heating time should not be more than 30 Sec
5. Treatment time should be between 20-60 sec
6. Visual and Audible cues to ensure working temperature reached
7. Beep sound should be there at 30, 45 and 60 sec
8. Built in timer should be there
9. Light source should be LED
10. Probe tip temperatures controlled to reach 100 degrees Celsius
11. Minimum 4 probes tip required and probes should not have any sharp edges.
Out of these four probes, three probes should be flat and one probe can have a gentle nipple extrusion not exceeding 5mm (so as to anchor in center of cervix but not to ablate endocervix)"
12. Probe-16mm and 19 mm should be there
13. Probe should be removable
14. Probes should have heat protection system
15. Machine Should be power operated by rechargeable Li-ion battery and should be provided with at least 1 spare standby battery along with its charger. (In total 2 batteries with the machine)
16. Handheld device - should be easily carried in a case
17. Probes should be autoclavable or can be sterilized using chemical
18. Electrical safety conforms to the standards for electrical safety. IEC 60601-1/IEC 60601-1-2 and IEC 62133.
19. Certifications: It should be USFDA/ European CE (4 digit notified) / CDSCO registration or CDSCO License/ Manufacturer should be ISO 13485 Certified.

20. Should have 2 yr Warranty and 8 yr CMC.



Dr. Reena
निदेशक आचार्य एवं विभागाध्यक्ष
Director Professor & HOD
प्रसूति एवं स्त्री रोग विभाग / Deptt. of Obst. & Gynae.
स.आ.के.ए. एवं श्रीमती सु.कु. अस्पताल, नई दिल्ली-1
LNMC & Smt. S.K. Hospital, New Delhi-1



28/7/24
डॉ. स्वाति अग्रवाल Dr. Swati Agrawal
आचार्य Professor
कीर्तिलाली आर (006589) DMIC No (00658)
प्रसूति एवं स्त्री रोग विभाग Deptt of Obs & Gynae
स.आ.के.ए. एवं श्रीमती सु.कु. अस्पताल नई दिल्ली-1
LNMC & Smt S.K Hospital New Delhi-1