File no: M/286/2025-STR /E-17664 भारत सरकार

Government of India स्वास्थ्य सेवा महानिदेशालय

Directorate General of Health Services स्वास्थ्य एवं परिवार कल्याण मंत्रालय

Ministry of Health & Family Welfare लेडी हर्डिंग मेडिकल कॉलेज एवं श्रीमती सुचेता कृपलानी अस्पताल

तहा हाइन माइकल कालज ६५ आनला पुत्रसा हुन्यता । Lady Hardinge Medical College & Smt. Sucheta Kriplani Hospital शहीद भगत सिंह मार्ग, नईदिल्ली– ११०००१

Shaheed Bhagat Singh Marg, New Delhi-110001 भंडारअनुभाग/Stores Section Dated- 26.05.2025

PUBLIC NOTICE

The LHMC & Smt. S. K. Hospital intent to procure the following items/ stores under GFR-155 immediately;

Sr. No.	Name of Item	Description	Quantity Required	Warranty	Remarks
1	Intraosseous Device	Specifications attached	06 Nos.	Specifications attached	_

The Competent Suppliers/ Service Providers are requested to submit/ drop their quotations (as per the format enclosed) within 02 days (or earlier) from the issue of this notice in the Tender Box kept in the

- 1. The quotations received after the prescribed dates will not be entertained in any condition.
- 2. The quotation should be addressed to the Chairperson Local Purchase Committee LHMC.
- 3. The quotation must bear Name, Contact Number and Address details of the bidder.

(Avanish Singh) Stores Officer LHMC & Smt. S.K. Hospital, New Delhi

To,

- Notice Board, Office of M.S, LHMC
- 2. Notice Board, Stores Section, LHMC
- Official Website of this Institution.

The duly sealed quotations must be having following information as under:

Quotation Format

Sr. No.		Rate Offered (In Rs.)	Units Packing Details	Quantity to supplied	be Amount Rupees
	Offered Make/ Brand:				

	Offered model:		1					
		1 00						
	,							
Total		Rs.						
GST F	Percentage	%						
GST A	Amount	Rs.						
Total	Amount (included G	Rs.						
Rupe	es in Words:				1			
/- (Inc	luded GST)							
Note:								
1.	 In case of discrepancy between rate offered (unit rate) and total amount written in figures, the rate offered (unit rate) shall prevail. 							
2.	2. In case of discrepancy between total amount written in figures & words, the amount mentione							

Apart from the above the quotations must bearing the following details as under;

Offered Warranty (if asked or applicable)

in the word shall prevail.

GST Number of the bidder:

Complete Mailing Address (with Pin Code):

Mobile Number:

Full Name of Owner (Please enclose the copy of PAN Card & Adhar Card):

Name of Father/ Husband:

PAN No.:

Aadhar No .:

e-mail address:

Name of Signatory should be there in the quotation

Instructions-

- In case of non-submission of any of the above details the quotation will be treated incomplete & rejected.
- 2. The demonstration of the quoted item must be provided if & when asked by the Institution, failing which the offer will be treated as rejected.

Digitally signed by AVANISH SINGH Date: 26-05-2025

15:22:56

Specifications:

- The device should be efficient conduit to establish intraosseous vascular access.
- The device should be one hand-held power driver with lithium battery, expected shelf life of 10 years or approximately 500 insertions.
- The device should have LED indicator for battery status indicating Green for sufficient power and Red blinker for less than 10% battery life.
- The device should be US FDA and CE certified with notified four digits
- The device's needles should be approved for Intraosseous access with depth marking and symmetrical cutting tip
- The device should include a stabilizer with adhesive to stabilize the needle in-situ, a luer-lock connector to connect to the universal infusion lines, a stylet and a safety cap.
- The device should be indicated to provide intraosseous access in the proximal tibia, distal tibia and proximal humerus of pediatric and adult patients
 - The device should be indicated to provide intraooseous access in the distal femur in pediatric patients
- The device should have IO needle sets, compatible with one hand-held power driver for insertion
- The device should have the needle sets that are Single use, sterile, 304 stainless steel with 15-gauge bore
- The device should be available in colour coded three sizes i.e. 15mm (with pink hub), 25 mm (with blue hub) & 45mm (yellow hub)
- The device's needle should be indicated to have dwelling time of 48 hours
- The driver should have minimum 500 thrusts for establishing IO access or ten years of shelf life.