F.No.M/1087/2024-STR/16081 / 968 4

भारत सरकार

Government of India स्वास्थ्य सेवा महानिदेशालय

Directorate General of Health Services स्वास्थ्य एवं परिवार कल्याण मंत्रालय

Ministry of Health & Family Welfare लेडी हर्डिंग मेडिकल कॉलेज एवं श्रीमती सुचेता कृपलानी अस्पताल

Lady Hardinge Medical College & Smt. Sucheta Kriplani Hospital

शहीद भगत सिंह मार्ग, नई दिल्ली– १९०००१

Shaheed Bhagat Singh Marg, New Delhi-110001

भंडार अनुभाग/Stores Section

Dated: 11/08/2025

PUBLIC NOTICE

The Director, LHMC & Associated Hospitals proposed to purchase under mentioned item for the Department of Obstt. & Gynae under GFR-166 (Single Tender Enquiry) i.e. on proprietary basis. The details of proposal under consideration are as under;

S.No.	Description	Details
1.	Name of item	Fetoscopy Set (Specifications Attached)
2.	Number of equipment to be procured	01 Set.
3.	Cost of equipment	Rs. 33,26,202.00/- GST extra as per applicable
4.	Warranty	Five years from the date of supply
5.	CMC after expiry of warranty	Five years
6.	Manufacturer/Supplier with	M/s Karl Storz SE & Co. KG. (Authorized Supplier) M/s Vishal Surgical Equipment Co. B-5, Rajendra Park, Pusa Road, New Delhi-110060.

The representations (if any) along with proper justification (supported by documents) should be submitted to the office of Director, LHMC and Associated Hospitals within 21 days, failing which, it will be presumed that the item is proprietary in nature and the proposal will be processed accordingly.

This issues with the approval of Competent Authority.

Digitally signed by AVANISH SINGH

Date: 11-08-2025 (Avanish Singh) 11:28:08 Stores Officer

11:28:08 Stores Officer LHMC & SSK Hospital, New Delhi

Copy to: Notice Board

Specification of Fetoscopy set

- Miniature Straight Forward Telescope 0 degree Set, curved, diameter 3.3mm, length 30 cm, with 30,000 pixels, 4 Fr. central working channel & 3 Fr. Lateral working channel QTY 1
- Miniature Straight Forward Telescope Odegree Set, straight, diameter 3.3mm, length 30 cm, with 30,000 pixels, 4 Fr. Central working channel & 3 Fr.lateral working channel QTY 1
- Miniature Straight Forward Telescope 0 degree, outer diameter: 1.3mm, length: 30.6 cm QTY 1
- Bipolar optical grasping forceps, size 3 mm, length 24.5 cm, serrated, flat jaws QTY 2
- Trocar, with pyramidal tip, drop-shaped profile, with luer-lock connector, size 3.6 X 5.4 mm, working length 10 cm QTY 2
- Plastic Container for Sterilization and Storage of Variable, Instrument Sets, perforated, with transparent lid, with silicone mat, single-level storage, (1 additional insert), external dimensions (w x d x h): 525 x 240 x 70 mm, including: Snap-in Clip. package of 12, Silicone Tie-Downs, package of 12, Tool - Qty 3
- 5 Years Warranty further 5 Years CMC

Note :- All items mentioned above should be USFDA/European CE with 4 digit notified body approved/BIS.

डा. मेना I Dr. Reena निदेशक आचार्या एवं विमागाच्यक्षा Director Professor & HOD प्रसुति एवं स्त्री रोग विचान / Deptt of Obst. & Gynae. ते.हा.ने.का. एवं श्रीमदी सु.कृ. अनिसाल, नुई दिल्ली-1 LHMC & Smt. S.H. Tospital New Delhi-1

Dr. JAYA CHAWLA MBBS, MS(DB GY), Fellow (Maternal Felat Medicine, AIIMS) Drichinson

Department of Observe - A Gynaecology ABVIMS & Dr RMt Hospital, New Delhi DARC 29/191

डा. शादान अला / Dr. Shadabami आचार्य / Professor डीएमसी सं.-30739 / DMC No.-30739 शब्द विशिष्टमा निमान / Depth of Surgery सबस्य व्य वश्येष क्राव्य, बात सरका / MOHSEW, Gov of loss ते.स.ने.का. व्य श्रीमती सु.क. अस्पताल, नई हिल्ली-11000 LHMC & Smt. S.K. Hospital, New Delit-11000

नेण I Dr. Shilpi Nain Member

Dr. Meenakshi 对 . Member I Dr. Manisha Kumar I Dr. Meenakshi 对 . Alember I Director Professor Dr. Meenakshi 对 . Alember I Director Professor Dr. 1899 I DMC No. 1899 I DMC Arran Dr. Meenakshi

ausle

5mt