

LIST OF NON-FUNCTIONAL EQUIPMENTS LHMC & SMT. S. K. HOSPITAL

S. No.	Name of Deptt.	Name of the equipment with Model & Serial No.	Total No. of equipment	Cost of equipment (Rs.), Warrantee & Maintenance	Date of Purchase	Date of Installation	Name of the supplier with contact detail	Date from which equipment is/are non-functional	Date from which equipment in under AMC/CMC/warranty if not reasons therefore	Reason for non-functioning of equipment if any	Details of part required for repair along with estimate expenditure	Action taken to repair the equipment & approximate time by which equipment will become functional	Action taken by M&E	Whether ripe for condemnation if yes from when and present book value
1	Anaesthesia	Steam Sterilizer Model: AMSCO 250L Make: Steris	01	\$ 35150.00 (including Carriage, Insurance & Freight) + Tax (including Cost of Indian Article) (Per unit) Warranty: 5 years from the date of installation + 5 years CMC 1st 40000 2nd 40000 3rd 40000 4th 40000 5th 40000	20-54/LHMC/Store/AT/Steam Sterilizer /2016/665 Dated: 24/09/2016	06/10/2017	M/s Steris India Pvt. Ltd.	07/08/2019	Under CMC	LCD Module Broken	LCD Module Needed Estimate : Rs. 493940.00	Repair Proposal received, recommended by the R/M Committee, the department has been asked for clarification regarding broken LCD Module	--	No
2	Biochemistry	ELISA Reader High End with Washer Model: Merilyzer Eliquant Quandra Make: M/s Meril	02	3090000.00 Each + VAT Warranty: 2 years from the date of installation + 5 years CMC/Unit 1st 110000 2nd 110000 3rd 110000 4th 110000 5th 110000	20-50/Store s/LHMC/Bio/HLL/2016-17/331 Dated : 02/07/2016	11/11/2016	M/s Meril Diagnostics Pvt. Ltd. Qty: 02 No.	January 2019	Under Warranty	Required multiple consumables	Multiple consumables Estimate : 139630.00	Supply order issued on 07/12/2019	--	No

		Diagnos tics Pvt. Ltd.												
3	Physiolog y	Exercise Physiolo gy System (Power Cube Ergo Cardio Pulmona ry System)	01	Euros 26290 (FOB) 3 Years Warrantee & 5 Years AMC 1 ST 34090 2 ND 37499 3 RD 41249 4 TH 45374 5 TH 49911 6 TH 54902 7 TH 60392	SO No. 70 Dated: 08/05/2 009	05/08/2 010	M/s Ganshor n Medizin Electron ic GmbH	July 2019	Under AMC	PCB Board malfunc tioned	Repair of Main PCB Board & Calibrati on Repair Estimate : Rs. 250000.0 0	Repair Order already issued on 24/09/20 19, The service provider has taken the componen t for repair to the OEM at Germany (outside of India)	--	No
4	Microbiol ogy	Biosafet y Cabinet Type-II Model: AC2-3E 1	01	US\$ 12122.00 Rs.816567.00 Warranty: 3 years from the date of installation + 5 years AMC/Unit 1 ST 14000 2 ND 15400 3 RD 16940 4 TH 18634 5 TH 20497	1041 dt. 23.12.20 13	10.6.201 4	M/s Krish Biomedic als	May 2019	Under AMC	UV Light, Blower Motor, Touch Keypad not working	Need UV Light, Blower Motor, Touch Keypad Estimate : 130763.00	Work Order already issued on 07/12/20 19	--	No